

Electroretinography 3 JAX_ERG_003

Purpose

Full-field electroretinogram (ERG) is a mass electrical response of the retina to a light stimulus. The ERG contains four components: a-wave, b-wave, c-wave and FO-wave. These four components reflect the responsiveness of retinal photoreceptor cells and other neurons as a measure of visual function.

Experimental Design

- **Minimum number of animals** : 4
- **Age at test:** Week 15
- **Sex:** We do not expect the results of this test to show sexual dimorphism

Equipment

Celeris - Diagnosys LLC

Procedure

Procedure

1. Transfer mice from the animal room to the testing room.
 2. Mice are dark adapted overnight (alternatively: minimum of 2 hours) for testing procedures.
 - a. All rod testing procedures are done under dim red light after a minimum 2-hour dark adaptation period
- Following dark adaptation one drop of an appropriate mydriatic solution is applied to each eye to induce pupil dilation.
 - Once the pupils have dilated mice are anesthetized with the inhalation of 2% Isoflurane.
 - When the mouse has reached the proper plane of anesthesia, the mouse is placed on a heated platform and the nose is placed in the nose cone for continuous anesthetic inhalation.
 - An appropriate ophthalmic lubricant is applied to the electrodes before placing on each cornea.
 - The scotopic test is performed with pulsing white light.
 - A 3-minute light-adaptation follows before the photopic test is run. At this time an additional drop of ophthalmic lubricant should be added to each eye without adjusting the electrodes.
 - After 3 minutes of light adaptation, photopic ERGs are obtained with brighter white flashes at varying stimulation intensities.

- Once the photopic test is complete, carefully remove the recording electrodes and apply a generous amount of an ophthalmic gel to both eyes and allow the mouse to recover in a clean heated pen until fully conscious.
- Return the mouse to its home pen.
- Clean the electrodes with sterile water followed by 70% ethanol.
- Save the test results and export to the server.

Notes

Amplitude and timing measures of the ERG waveform are taken:

1. The a-wave amplitude is measured from the pre-stimulus baseline to the lowest negative trough; the b-wave amplitude is measured from the trough of the a-wave to the following highest peak; the c-wave amplitude is measured from the pre-stimulus baseline to the following highest peak; the FO- wave amplitude is measured from the peak of the c-wave to the subsequent lowest trough.
2. Implicit time (t) is measured from stimulus onset to the trough or peak of each wave.

Parameters and Metadata

R-blind JAX_ERG_001_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Options: Yes, No,

L-blind JAX_ERG_002_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Options: Yes, No,

RE-a (uV) [Scotopic] JAX_ERG_003_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

RE-b (uV) [Scotopic] JAX_ERG_004_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

RE-c (uV) [Scotopic] JAX_ERG_005_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

LE-a (uV) [Scotopic] JAX_ERG_006_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

LE-b (uV) [Scotopic] JAX_ERG_007_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

LE-c (uV) [Scotopic] JAX_ERG_008_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

RE-a (uV) [Photopic] JAX_ERG_009_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

RE-b (uV) [Photopic] JAX_ERG_010_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

LE-a (uV) [Photopic] JAX_ERG_011_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

LE-b (uV) [Photopic] JAX_ERG_012_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

RE-a (ms) [Scotopic] JAX_ERG_013_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

RE-b (ms) [Scotopic] JAX_ERG_014_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

RE-c (ms) [Scotopic] JAX_ERG_015_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

LE-a (ms) [Scotopic] JAX_ERG_016_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

LE-b (ms) [Scotopic] JAX_ERG_017_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

LE-c (ms) [Scotopic] JAX_ERG_018_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

RE-a (ms) [Photopic] JAX_ERG_019_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

RE-b (ms) [Photopic] JAX_ERG_020_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

LE-a (ms) [Photopic] JAX_ERG_021_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

LE-b (ms) [Photopic] JAX_ERG_022_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

RE-FO-like (uV) [Scotopic] JAX_ERG_023_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

LE-FO-like (uV) [Scotopic] JAX_ERG_024_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: uV

RE-FO-like (ms) [Scotopic] JAX_ERG_025_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

LE-FO-like (ms) [Scotopic] JAX_ERG_026_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

Comments JAX_ERG_027_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Fundus file JAX_ERG_028_001 | v1.0

seriesMediaParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Experimenter ID JAX_ERG_029_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: experimenter_id

Stimulus protocol JAX_ERG_030_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Topical agents JAX_ERG_031_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Software version JAX_ERG_032_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

SOP version JAX_ERG_033_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Stimulator JAX_ERG_034_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

C-wave (cd.s/m²) JAX_ERG_035_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Photopic ERG (cd.s/m²) JAX_ERG_036_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false
