

# Hematology IMPC\_HEM\_002

## Purpose

Hematological assessment of blood determines blood cell counts (white blood cells, red blood cells, hemoglobin, and platelets) and additional hematological parameters (hematocrit, mean cell volume, mean corpuscular hemoglobin, mean cell hemoglobin concentration) can be derived using these indices. These tests will indicate abnormalities in the production of blood and its components (blood cells and hemoglobin) as well as in the associated blood-forming organs.

Ontological description: MP:0002429 - abnormal blood cell morphology/development.

## Experimental Design

- **Minimum number of animals** : 7M + 7F
- **Age at test**: Week 16
- **Sex**: We would expect the results of this test to show sexual dimorphism

## Equipment

1. Hematology automated analyzers (e.g. Beckman Coulter AcT Diff , Siemens Advia 2120 or Hemavet Multispecies Hematology Analyzer HV950FS Drew Scientific, CT, U. S.A.)
2. Rotary agitator

## Procedure

Set up the hematological analyser and perform QC analyses of the control reagents in accordance with the guidelines provided by the manufacturer.

### Sample collection and preparation:

- a. Collect the appropriate volume of blood required for the hematology analyser being used for assessment (~200µl), in an EDTA coated tube with the relevant blood collection procedure (see IMPC protocol Blood collection by retro-orbital puncture). The time of day for collection is in the morning, starting no earlier than 07:30.
- b. Mix the blood sample on a rotary mixer immediately following collection for a minimum of 30 minutes and keep the sample at room temperature (for no more than 2 hours) pending analysis. Samples must *not* be frozen at this stage.

- c. Analysis of samples is optimally done on the day of collection. When not possible the blood samples can be stored at 2-8°C for up to 24 hours. Long term storage of whole blood is not recommended. All samples are allowed to come to room temperature prior to analysis.

**Analysis:**

- a. Perform hematological assessment of each sample including: white and red blood cell counts, hemoglobin and platelets in accordance with the analyser being used.
- b. Derive additional parameters for the sample that may be estimated from the initial assessment such as: hematocrit, mean cell volume, mean corpuscular hemoglobin and mean cell hemoglobin concentration.

**Notes**

Blood collection for Clinical Chemistry and Hematology is usually performed as a non-fasting, terminal procedure but can be performed as a non-terminal procedure under certain circumstances. Mice from the terminal procedure may be used for subsequent gross pathology and other procedures included in terminal assessments. Whole blood (for Hematology) and plasma (for Clinical Chemistry) require different collection tubes so two independent samples are required from each mouse. Dilution of blood is highly discouraged, but is allowed when the total necessary amount is not obtained. If dilution is necessary then the assays should be done in one run.

The information about the date of the experiment, that is the date when the measurement is performed, is an important parameter which is to be submitted in the Experiment xml file (dateOfExperiment="2013-02-28").

**Data QC**

- 1. Sample must be free of blood clots in order to be analyzed.
- 2. Some results from hemolysed samples should not be reported.
- 3. Perform routinely and immediately prior to sample analysis:
  - a. assessment of control samples with different levels of hematology phenotypes (abnormally low; normal; abnormally high).
  - b. analysis of the graphical reports generated for each control level to ensure that they lie within their respective ranges.

**Metadata and examples**

Metadata	Example
Equipment ID	ID of the machine used when more than 1 is used having same model and manufacturer. E.g. machine 1, machine 2, machine Minnie, machine Mickey Mouse, etc.

Equipment manufacturer	Manufacturer of the equipment. E.g. SIEMENS.
Equipment model	Model of the equipment. E.g. ADVIA120.
Blood collection tubes	The tubes used for blood collection. E.g. Sarstedt Li-Heparin gel tubes or Kabe Labortechnik Lithium heparin coated tubes.
Method of blood collection	Concise description of the method used for blood collection. E.g. Retro-orbital puncture.
Anesthesia used for blood collection	The drug used for anaesthesia during blood collection. E. g. Isoflurane.
Anticoagulant	Anticoagulant drug used for blood collection. E.g. EDTA.
Samples kept on ice between collection and analysis?	Yes/No
Storage temperature from blood collection till measurement	E.g. 2°C
Date and time of blood collection	Time of day for collection is in the morning, starting no earlier than 07:30. E.g. Year, month, day, time.
Date of measurement	The day of blood analysis. E.g. Year, month, day.
ID for blood collection SOP	ID of the protocol followed for blood collection. Can be a center specific protocol. E.g. ESLIM_024_001
	The chip card contains the settings and thresholds that are used to calculate the numbers of cell types in a blood sample. As the blood cell sizes differ between the species, there are different thresholds for the

Chip card	<p>categorization and therefore there are different chip cards for different species (mouse strains). Eg. C57BL/6 chip card.</p> <p>The chip cards really look like a chip card. You put them into a slot on the haematology device and then you start measuring the haematological parameters of the corresponding blood samples.</p>
Blood collection experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures indicating the experimenter who collected the blood. E.g. Harw_001, or 1/2/3.
Blood analysis experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures indicating the experimenter who analyzed the blood. E.g. Harw_001, or 1/2/3.
Date equipment last calibrated	Most recent date in which the equipment (or any part of) used in the procedure was subject to a calibration event.
Date and time of sacrifice	The date and time when the mouse is sacrificed.

## Parameters and Metadata

### White blood cell count IMPC\_HEM\_001\_001 | v1.3

simpleParameter

**Req. Analysis:** false

**Req. Upload:** true

**Is Annotated:** true

**Unit Measured:** 10<sup>3</sup>/ul

**Description:** white\_blood\_cell\_count

---

## Red blood cell count IMPC\_HEM\_002\_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured:  $10^6/\text{ul}$

Description: red\_blood\_cell\_count

---

## Hemoglobin IMPC\_HEM\_003\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: g/dl

Description: hemoglobin

---

## Hematocrit IMPC\_HEM\_004\_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: %

Description: hematocrit

---

## Mean cell volume IMPC\_HEM\_005\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: fL

Description: mean\_cell\_volume

---

## Mean corpuscular hemoglobin IMPC\_HEM\_006\_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: pg

Description: mean\_corpuscular\_hemoglobin

---

## Mean cell hemoglobin concentration IMPC\_HEM\_007\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: g/dl

Description: mean\_cell\_hemoglobin\_concentration

---

## Platelet count IMPC\_HEM\_008\_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: 10<sup>3</sup>/ul

Description: platelet\_count

---

## Equipment ID IMPC\_HEM\_009\_001 | v1.1

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Description: equipment\_name

---

## Equipment manufacturer IMPC\_HEM\_010\_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Description: equipment\_manufacturer

**Options:** Scil animal care company Gmbh, Drew Scientific Instrument, Beckman Coulter, Siemens Medical Solutions Diagnostics, Siemens Healthcare Diagnostics Ltd, Sysmex Deutschland GmbH, Abbot Laboratories, Mindray, IDEXX,

---

## Equipment model IMPC\_HEM\_011\_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

**Description:** equipment\_model

**Options:** Advia 120, Advia 2120, Scil Vet abc, Hemavet 950 FS, Ac-T diff Analyzer, XT-2000iV, CELL-DYN 3700, Scil Vet abc Plus+, BC-5300 Vet, ProCyte Dx, Advia 2120i,

---

## Anesthesia used for blood collection IMPC\_HEM\_012\_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

**Description:** anesthesia\_used\_for\_blood\_collection

**Options:** Gas anaesthesia with Isoflurane,  
Injection narcosis with Ketamine (100mg/kg)/Xylazine (10mg/kg),  
Injection narcosis with Ketamine (100mg/kg)/Xylazine (10mg/kg)/Antipamezole (Antisedan, 1mg/kg),  
Injection narcosis with Ketamine (110mg/kg)/Xylazine (11mg/kg),  
Injection narcosis with Ketamine (110mg/kg)/Xylazine (11mg/kg)/ Antipamezole (Antisedan, 1mg/kg),  
No anesthesia, Injection narcosis with Ketamine (137mg/kg)/Xylazine (6.6mg/kg),  
Injection narcosis with Tribromoethanol (Avertin),

---



## Method of blood collection IMPC\_HEM\_013\_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Description: method\_of\_blood\_collection

Options: Cardiac puncture, Retro-orbital puncture, Tail vein, Saphenous vein,

---

## Anticoagulant IMPC\_HEM\_014\_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: anticoagulant

Options: EDTA, K(1)-EDTA, K(2)-EDTA, K(3)-EDTA, No,

---

## Samples kept on ice between collection and analysis IMPC\_

HEM\_018\_001 | v1.2

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Description: samples\_kept\_on\_ice\_between\_collection\_and\_analysis\_

Options: Yes, No,

---

## ID for blood collection SOP IMPC\_HEM\_020\_001 | v1.1

procedureMetadata

**Req. Analysis:** false

**Req. Upload:** true

**Is Annotated:** false

**Description:** id\_for\_blood\_collection\_sop

**Options:** ESLIM\_024\_001, sop.inv.019, RIKENMPP\_003a\_003, PHENO\_CBC, sop.inv.063, CCP-Hemo\_SOP,

---

## Date and time of blood collection IMPC\_HEM\_021\_001 | v1.2

procedureMetadata

**Req. Analysis:** false

**Req. Upload:** true

**Is Annotated:** false

**Description:** date\_and\_time\_of\_blood\_collection

---

## Chip card number IMPC\_HEM\_023\_001 | v1.1

procedureMetadata

**Req. Analysis:** true

**Req. Upload:** false

**Is Annotated:** false

**Description:** chip\_card\_number

**Options:** C57/BL6 chip card, Mouse Card (E0510051710), Mouse Card (E0401091230), No chip card, No chip card (Advia analyser),

---

## Blood collection experimenter ID IMPC\_HEM\_024\_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

---

## Date equipment last calibrated IMPC\_HEM\_025\_001 | v1.2

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

---

## Storage temperature from blood collection until measurement IMPC\_HEM\_026\_001 | v1.3

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Unit Measured: C

Options: 22, 4, 25, 18-22, 23,

---

## Blood collection tubes IMPC\_HEM\_015\_001 | v1.2

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

**Options:** Kabe Labortechnik 200ul EDTA, Kabe Labortechnik 1ml EDTA,  
Drummond EDTA Microcaps, Microvette 500 K3E, Eppendorf 1.7ml,

---

## Date and time of sacrifice IMPC\_HEM\_016\_001 | v1.3

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

## Blood analysis experimenter ID IMPC\_HEM\_017\_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

## Mean platelet volume IMPC\_HEM\_019\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: fL

---

## Red blood cell distribution width IMPC\_HEM\_027\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: %

---

## Fight wounds IMPC\_HEM\_028\_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Options: Yes, No,

---

## Neutrophil differential count IMPC\_HEM\_029\_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: %

### Description:

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

## Neutrophil cell count IMPC\_HEM\_030\_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured:  $10^3/\text{ul}$

### Description:

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

## Lymphocyte differential count IMPC\_HEM\_031\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: %

### Description:

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

## Lymphocyte cell count IMPC\_HEM\_032\_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

**Unit Measured:** 10<sup>3</sup>/ul

**Description:**

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

**Monocyte differential count** IMPC\_HEM\_033\_001 | v1.2

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** %

**Description:**

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

**Monocyte cell count** IMPC\_HEM\_034\_001 | v1.3

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** 10<sup>3</sup>/ul

**Description:**

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

**Eosinophil differential count** IMPC\_HEM\_035\_001 | v1.2

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** %

**Description:**

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

**Eosinophil cell count** IMPC\_HEM\_036\_001 | v1.3

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** 10<sup>3</sup>/ul



**Description:**

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

**Basophil cell count** IMPC\_HEM\_037\_001 | v1.1

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** 10<sup>3</sup>/ul

**Description:**

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

**Basophil differential count** IMPC\_HEM\_038\_001 | v1.0

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** %

**Description:**

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

**Large Unstained Cell (LUC) count** IMPC\_HEM\_039\_001 | v1.0

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** 10<sup>3</sup>/ul

**Description:**

A white blood cell (WBC) count measures the number of white blood cells in your blood. A WBC differential determines the percentage of each type of white blood cell present in your blood. A differential can also detect immature white blood cells or any abnormalities, both of which are signs of a potential problem.

---

**Large Unstained Cell (LUC) differential count** IMPC\_HEM\_040\_0

01 | v1.0

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** %

---

## Sample clotted IMPC\_HEM\_041\_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Options: Yes, No,

---

## Service-related calibration start date IMPC\_HEM\_042\_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: false

Is Annotated: false

### Description:

Harwell-required metadata parameter (req analysis) due to ADVIA analyser causing shift in data.

---

## LIH (Hemolysis Severity - available on AU analysers) IMPC\_HEM\_043\_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

### Description:

Copied from the same parameter in Clinical Blood Chemistry, as the same blood samples are used for both procedures, and the level of hemolysis has an effect on the results of hematology as well as CBC

---

